

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081722\
 Data File : VL039550.D
 Acq On : 17 Aug 2022 16:16
 Operator : SY/AP
 Sample : N4125-01 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 08/18/2022
 Supervised By : Mahesh Dadoda 08/19/2022

Quant Time: Aug 18 05:38:32 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 17 12:06:24 2022
 QLast Update : Wed Aug 17 12:06:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	1127592	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.661	114	2885891	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.475	117	2611367	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	1965845	10.000	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.000%
Target Compounds						
9) Propene	2.723	41	19540	0.317	ppbv	93
15) Acetone	3.504	43	119172	0.935	ppbv	92
20) Methylene Chloride	3.954	84	93726	1.807	ppbv	96
26) Hexane	5.230	57	47170	0.423	ppbv #	39
30) Cyclohexane	6.613	84	11751m	0.127	ppbv	
53) Toluene	9.230	91	53089m	0.215	ppbv	
59) m/p-Xylene	12.352	91	69639m	0.256	ppbv	
74) 1,2,4-Trimethylbenzene	16.108	105	67230m	0.216	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081722\
Data File : VL039550.D
Acq On : 17 Aug 2022 16:16
Operator : SY/AP
Sample : N4125-01 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/18/2022
Supervised By :Mahesh Dadoda 08/19/2022

Quant Time: Aug 18 05:38:32 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Aug
QLast Update : Wed Aug 17 12:06:24 2022
Response via : Initial Calibration

