

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081722\
 Data File : VL039555.D
 Acq On : 17 Aug 2022 19:36
 Operator : SY/AP
 Sample : N4125-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/18/2022
 Supervised By :Mahesh Dadoda 08/19/2022

Quant Time: Aug 18 05:39:42 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Aug :
 QLast Update : Wed Aug 17 12:06:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.208	49	1025631	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.664	114	2651284	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.475	117	2398215	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.783	95	1818087	10.070	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.700%
Target Compounds						
						Qvalue
13) Ethanol	3.272	45	15095m	3.746	ppbv	
15) Acetone	3.504	43	68679	0.593	ppbv	95
19) Isopropyl Alcohol	3.616	45	7275m	0.128	ppbv	
20) Methylene Chloride	3.954	84	62006	1.314	ppbv	95
26) Hexane	5.227	57	32024	0.316	ppbv #	42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081722\
Data File : VL039555.D
Acq On : 17 Aug 2022 19:36
Operator : SY/AP
Sample : N4125-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/18/2022
Supervised By :Mahesh Dadoda 08/19/2022

Quant Time: Aug 18 05:39:42 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Aug
QLast Update : Wed Aug 17 12:06:24 2022
Response via : Initial Calibration

