

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081822\
 Data File : VL039561.D
 Acq On : 18 Aug 2022 16:21
 Operator : SY/AP
 Sample : M4526-01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :

Quant Time: Aug 19 03:02:49 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 17 12:06:24 2022
 QLast Update : Wed Aug 17 12:06:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.208	49	997621	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.668	114	2612454	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.478	117	2379521	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.786	95	1831826	10.226	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	= 102.300%	
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.944	62	1764297	28.618	ppbv	96
10) Heptane	7.594	43	4356513	34.434	ppbv	100
12) 1,1,2-Trichlorotrifluo...	4.089	101	2667940	24.520	ppbv	97
24) 1,1-Dichloroethane	4.584	63	3301287	33.103	ppbv	98
26) Hexane	5.234	57	4174392	42.358	ppbv #	34
30) Cyclohexane	6.619	84	2659202	32.496	ppbv #	100
36) Benzene	6.388	78	4767969	23.947	ppbv	97
48) Dibromochloromethane	9.719	129	2439027	19.184	ppbv	98
53) Toluene	9.233	91	5403192	24.221	ppbv	98
54) 1,2-Dibromoethane	10.021	107	4530225	46.442	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081822\
Data File : VL039561.D
Acq On : 18 Aug 2022 16:21
Operator : SY/AP
Sample : M4526-01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :

Quant Time: Aug 19 03:02:49 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 17 12:06:24 2022
QLast Update : Wed Aug 17 12:06:24 2022
Response via : Initial Calibration

