

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081822\
 Data File : VL039570.D
 Acq On : 18 Aug 2022 22:24
 Operator : SY/AP
 Sample : N3980-13 0.03 LOD
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT3-2022

Quant Time: Aug 19 03:06:29 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 17 12:06:24 2022
 QLast Update : Wed Aug 17 12:06:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	913696	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.664	114	2448413	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	2178417	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	1701717	10.377	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.800%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.944	62	1919m	0.034	ppbv	
52) Tetrachloroethene	10.629	164	2331m	0.032	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.060	83	4051m	0.037	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081822\
Data File : VL039570.D
Acq On : 18 Aug 2022 22:24
Operator : SY/AP
Sample : N3980-13 0.03 LOD
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT3-2022

Quant Time: Aug 19 03:06:29 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 17 12:06:24 2022
QLast Update : Wed Aug 17 12:06:24 2022
Response via : Initial Calibration

