

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL082018\
 Data File : VL032415.D
 Acq On : 20 Aug 2018 22:01
 Operator : SY/AP
 Sample : J4472-02DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-2DL

Manual Integrations
 APPROVED

MMDadoda
 8/22/2018 10:50:37 AM

Quant Time: Aug 21 03:05:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
 QLast Update : Wed Aug 15 14:58:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.81	49	1684046	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.35	114	3228495	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.30	117	3259614	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.64	95	2449228	9.85	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.07	41	10560	0.16	ppbv	92
11) Trichlorofluoromethane	4.05	101	16823m	0.09	ppbv	
13) Ethanol	3.71	45	16955m	0.99	ppbv	
15) Acetone	3.96	43	95211	0.60	ppbv	96
19) Isopropyl Alcohol	4.09	45	43178	0.40	ppbv #	85
20) Methylene Chloride	4.45	84	14008	0.23	ppbv	86
76) 1,4-Dichlorobenzene	17.37	146	165208m	0.78	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL082018\
Data File : VL032415.D
Acq On : 20 Aug 2018 22:01
Operator : SY/AP
Sample : J4472-02DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-2DL

Manual Integrations
APPROVED

MMDadoda
8/22/2018 10:50:37 AM

Quant Time: Aug 21 03:05:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 15 14:58:02 2018
Response via : Initial Calibration

