

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL082020\
 Data File : VL035471.D
 Acq On : 20 Aug 2020 10:33
 Operator : SY/AP
 Sample : VL0820ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0820ABL01

Manual Integrations
 APPROVED

MMDadoda
 8/21/2020 11:24:04 AM

Quant Time: Aug 21 06:14:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 1
 QLast Update : Mon Aug 17 19:28:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1109371	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4894235	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.09	117	4851256	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	3446157	10.15	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
20) Methylene Chloride	4.35	84	12282m	0.119	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL082020\
Data File : VL035471.D
Acq On : 20 Aug 2020 10:33
Operator : SY/AP
Sample : VL0820ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0820ABL01

Manual Integrations
APPROVED

MMDadoda
8/21/2020 11:24:04 AM

Quant Time: Aug 21 06:14:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug
QLast Update : Mon Aug 17 19:28:30 2020
Response via : Initial Calibration

