

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL082224\
 Data File : VL041586.D
 Acq On : 22 Aug 2024 14:14
 Operator : SY/MD
 Sample : P3710-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CD132DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/23/2024
 Supervised By :Mahesh Dadoda 08/23/2024

Quant Time: Aug 23 01:32:19 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 23 15:15:50 2024

QLast Update : Fri Aug 23 01:15:26 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	889754	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.652	114	2364531	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	2147832	10.000	ppbv	# 0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	1612195	9.754	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.500%
Target Compounds						
					Qvalue	
9) Propene	2.722	41	19059	0.392	ppbv #	82
13) Ethanol	3.259	45	266593m	61.301	ppbv	
15) Acetone	3.484	43	521292	5.156	ppbv	98
19) Isopropyl Alcohol	3.591	45	257406	4.355	ppbv #	94
20) Methylene Chloride	3.947	84	39791	1.056	ppbv	91
25) Ethyl Acetate	5.217	43	48825m	0.243	ppbv	
34) 2-Butanone	4.803	43	20347	0.178	ppbv #	86
52) Tetrachloroethene	10.606	164	10862m	0.163	ppbv	
53) Toluene	9.217	91	79420	0.413	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL082224\
Data File : VL041586.D
Acq On : 22 Aug 2024 14:14
Operator : SY/MD
Sample : P3710-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CD132DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/23/2024
Supervised By :Mahesh Dadoda 08/23/2024

Quant Time: Aug 23 01:32:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Aug 23 01:15:26 2024
Response via : Initial Calibration

