

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL082320\
 Data File : VL035495.D
 Acq On : 23 Aug 2020 10:45
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Quant Time: Aug 23 12:02:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sun Aug 23 10:33:49 2020
 QLast Update : Sun Aug 23 10:33:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1019235	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4470717	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	4500400	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	3256742	10.34	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.25	62	8048	0.101	ppbv #	86
32) 1,1,1-Trichloroethane	6.51	97	22182	0.091	ppbv	97
35) Carbon Tetrachloride	7.01	117	21226m	0.089	ppbv	
38) Trichloroethene	7.84	130	18016	0.099	ppbv #	88
52) Tetrachloroethene	11.23	164	18346	0.092	ppbv	92
63) 1,1,2,2-Tetrachloroethane	13.70	83	30510	0.095	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL082320\
Data File : VL035495.D
Acq On : 23 Aug 2020 10:45
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Quant Time: Aug 23 12:02:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sun Aug 23 10:33:49 2020
QLast Update : Sun Aug 23 10:33:49 2020
Response via : Initial Calibration

