

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL082320\
 Data File : VL035502.D
 Acq On : 23 Aug 2020 17:19
 Operator : SY/AP
 Sample : L3510-01DL 4X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E053-1100DL

Quant Time: Aug 24 02:28:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Sun Aug 23 12:40:32 2020
 QLast Update : Sun Aug 23 12:40:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	988059	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4306757	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	4301760	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	3138640	10.21	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	3.47	94	668511	10.641	ppbv	98
18) 1,1-Dichloroethene	4.29	96	704107	8.134	ppbv	97
29) Chloroform	5.75	83	1410664	6.066	ppbv	99
38) Trichloroethene	7.84	130	1509019	9.641	ppbv	98
42) Bromodichloromethane	7.80	83	2545998	11.631	ppbv	99
47) 1,1,2-Trichloroethane	9.49	97	1054503	7.440	ppbv	98
48) Dibromochloromethane	10.32	129	1876813	8.319	ppbv	100
58) Ethyl Benzene	12.73	91	2870318	6.215	ppbv	97
76) 1,4-Dichlorobenzene	17.16	146	3709805	9.889	ppbv	100
77) 1,2-Dichlorobenzene	17.84	146	3538974	9.574	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL082320\
Data File : VL035502.D
Acq On : 23 Aug 2020 17:19
Operator : SY/AP
Sample : L3510-01DL 4X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E053-1100DL

Quant Time: Aug 24 02:28:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sun Aug 23 12:40:32 2020
QLast Update : Sun Aug 23 12:40:32 2020
Response via : Initial Calibration

