

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL082720\
 Data File : VL035515.D
 Acq On : 27 Aug 2020 10:33
 Operator : SY/AP
 Sample : L3510-01DL 4X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E053-1100DL

Quant Time: Aug 28 11:12:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Aug 27 06:46:28 2020
 QLast Update : Thu Aug 27 06:46:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1013860	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	4334416	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	4344464	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	3173232	10.29	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	3.48	94	670365	10.776	ppbv	100
18) 1,1-Dichloroethene	4.30	96	701599	8.238	ppbv	98
29) Chloroform	5.76	83	1421256	6.232	ppbv	100
38) Trichloroethene	7.86	130	1517688	9.312	ppbv	98
42) Bromodichloromethane	7.81	83	2549660	11.787	ppbv	99
47) 1,1,2-Trichloroethane	9.50	97	1011312	7.231	ppbv	99
48) Dibromochloromethane	10.33	129	1818143	8.260	ppbv	100
58) Ethyl Benzene	12.74	91	2729105	6.092	ppbv	95
76) 1,4-Dichlorobenzene	17.18	146	3044401	8.560	ppbv	98
77) 1,2-Dichlorobenzene	17.86	146	2771663	7.933	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL082720\
Data File : VL035515.D
Acq On : 27 Aug 2020 10:33
Operator : SY/AP
Sample : L3510-01DL 4X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E053-1100DL

Quant Time: Aug 28 11:12:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Aug 27 06:46:28 2020
QLast Update : Thu Aug 27 06:46:28 2020
Response via : Initial Calibration

