

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL082720\
 Data File : VL035516.D
 Acq On : 27 Aug 2020 11:15
 Operator : SY/AP
 Sample : L3510-01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E053-1100

Quant Time: Aug 28 11:13:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Aug 27 06:46:28 2020
 QLast Update : Thu Aug 27 06:46:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	955980	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	4082032	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	4134682	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	3119229	10.62	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	3.48	94	2412652	41.130	ppbv	100
18) 1,1-Dichloroethene	4.30	96	2577434	32.098	ppbv	95
29) Chloroform	5.76	83	4785718	22.256	ppbv	95
38) Trichloroethene	7.86	130	4991003	32.517	ppbv	97
42) Bromodichloromethane	7.82	83	7726820	37.929	ppbv #	86
47) 1,1,2-Trichloroethane	9.50	97	3642959	27.658	ppbv	97
48) Dibromochloromethane	10.34	129	6311199	30.445	ppbv	97
58) Ethyl Benzene	12.75	91	8578639	20.120	ppbv #	86
76) 1,4-Dichlorobenzene	17.18	146	9288015	27.440	ppbv	91
77) 1,2-Dichlorobenzene	17.86	146	8921071	26.828	ppbv	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL082720\
Data File : VL035516.D
Acq On : 27 Aug 2020 11:15
Operator : SY/AP
Sample : L3510-01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E053-1100

Quant Time: Aug 28 11:13:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Aug 27 06:46:28 2020
QLast Update : Thu Aug 27 06:46:28 2020
Response via : Initial Calibration

