

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL083122\
 Data File : VL039606.D
 Acq On : 31 Aug 2022 13:24
 Operator : SY/AP
 Sample : N4408-02DL 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVE-EFF-082222DL

Quant Time: Aug 31 23:44:50 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 31 2022
 QLast Update : Fri Aug 26 06:05:16 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/01/2022
 Supervised By : Mahesh Dadoda 09/02/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.214	49	1017794	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.674	114	2622378	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.484	117	2286545	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.793	95	1654408	9.591	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.900%
Target Compounds						Qvalue
15) Acetone	3.497	43	246378	2.161	ppbv	93
20) Methylene Chloride	3.960	84	402762	8.695	ppbv	92
26) Hexane	5.237	57	187557	1.884	ppbv #	41
31) cis-1,2-Dichloroethene	5.108	61	102991	1.068	ppbv	89
32) 1,1,1-Trichloroethane	6.025	97	21467m	0.157	ppbv	
34) 2-Butanone	4.819	43	46080m	0.353	ppbv	
38) Trichloroethene	7.311	130	299645	2.798	ppbv	96
52) Tetrachloroethene	10.629	164	79175m	1.047	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL083122\
Data File : VL039606.D
Acq On : 31 Aug 2022 13:24
Operator : SY/AP
Sample : N4408-02DL 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVE-EFF-082222DL

Quant Time: Aug 31 23:44:50 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 26 06:05:16 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 09/01/2022
Supervised By : Mahesh Dadoda 09/02/2022

