

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL090518\
 Data File : VL032454.D
 Acq On : 5 Sep 2018 12:43
 Operator : SY/AP
 Sample : VL0905ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0905ABL01

Manual Integrations
 APPROVED

MMDadoda
 9/7/2018 10:45:48 AM

Quant Time: Sep 06 15:18:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL083118AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 31 21:33:46 2018
 QLast Update : Fri Aug 31 21:33:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.80	49	1604821	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.33	114	3243906	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.29	117	3121201	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.63	95	2364737	10.20	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.95	43	55635m	0.38	ppbv	
20) Methylene Chloride	4.45	84	15207m	0.28	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL090518\
Data File : VL032454.D
Acq On : 5 Sep 2018 12:43
Operator : SY/AP
Sample : VL0905ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0905ABL01

Manual Integrations
APPROVED

MMDadoda
9/7/2018 10:45:48 AM

Quant Time: Sep 06 15:18:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL083118AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 31 21:33:46 2018
Response via : Initial Calibration

