

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090523\
 Data File : VL040662.D
 Acq On : 6 Sep 2023 1:09
 Operator : SY/MD
 Sample : 04281-05DL 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/06/2023
 Supervised By : Mahesh Dadoda 09/06/2023

Quant Time: Sep 06 02:09:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Aug 16 05:43:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	740648	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	2024882	10.000	ppbv	-0.01
55) Chlorobenzene-d5	11.468	117	1775964	10.000	ppbv	-0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	1439755	10.193	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.900%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.719	85	22185	0.139	ppbv	95
3) Chlorodifluoromethane	2.665	51	8821m	0.073	ppbv	
4) Chloromethane	2.800	50	7573m	0.160	ppbv	
9) Propene	2.697	41	5435	0.144	ppbv	84
11) Trichlorofluoromethane	3.552	101	11116	0.086	ppbv	95
13) Ethanol	3.240	45	9315m	0.886	ppbv	
15) Acetone	3.468	43	127636	1.409	ppbv	92
20) Methylene Chloride	3.925	84	53810	1.215	ppbv	95
26) Hexane	5.208	57	32564m	0.396	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090523\
Data File : VL040662.D
Acq On : 6 Sep 2023 1:09
Operator : SY/MD
Sample : 04281-05DL 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/06/2023
Supervised By :Mahesh Dadoda 09/06/2023

Quant Time: Sep 06 02:09:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Aug 16 05:43:50 2023
Response via : Initial Calibration

