

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090523\
 Data File : VL040665.D
 Acq On : 6 Sep 2023 9:21
 Operator : SY/MD
 Sample : 04261-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Quant Time: Sep 06 11:35:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Aug 16 05:43:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	899837	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	2381476	10.000	ppbv	-0.01
55) Chlorobenzene-d5	11.468	117	2156596	10.000	ppbv	-0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	1780152	10.379	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.800%
Target Compounds						
					Qvalue	
13) Ethanol	3.237	45	44021	3.446	ppbv	97
15) Acetone	3.462	43	192921	1.753	ppbv	98
20) Methylene Chloride	3.925	84	405433	7.538	ppbv	95
26) Hexane	5.208	57	196422	1.964	ppbv #	38

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090523\
Data File : VL040665.D
Acq On : 6 Sep 2023 9:21
Operator : SY/MD
Sample : 04261-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Quant Time: Sep 06 11:35:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Aug 16 05:43:50 2023
Response via : Initial Calibration

