

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090723\
 Data File : VL040676.D
 Acq On : 7 Sep 2023 20:11
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Sep 08 03:31:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Aug 16 05:43:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	881791	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.655	114	2273692	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	2022349	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.783	95	1579919	9.823	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.200%
Target Compounds						
					Qvalue	
15) Acetone	3.491	43	8101	0.075	ppbv #	47
20) Methylene Chloride	3.928	84	19336	0.367	ppbv	90
26) Hexane	5.211	57	3671	0.037	ppbv #	1
39) 1,2-Dichloropropane	6.655	63	578182	8.402	ppbv #	26

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090723\
Data File : VL040676.D
Acq On : 7 Sep 2023 20:11
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Sep 08 03:31:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Aug 16 05:43:50 2023
Response via : Initial Calibration

