

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090723\
 Data File : VL040680.D
 Acq On : 8 Sep 2023 9:51
 Operator : SY/MD
 Sample : 04323-01DL2 600X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/11/2023
 Supervised By : Mahesh Dadoda 09/11/2023

Quant Time: Sep 09 00:09:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Aug 16 05:43:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	789665	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.652	114	2093550	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	1891503	10.000	ppbv	-0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	1482509	9.855	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.600%
Target Compounds						
						Qvalue
15) Acetone	3.488	43	18929	0.196	ppbv	93
20) Methylene Chloride	3.928	84	19813	0.420	ppbv	92
26) Hexane	5.205	57	8361m	0.095	ppbv	
31) cis-1,2-Dichloroethene	5.076	61	5801m	0.073	ppbv	
38) Trichloroethene	7.288	130	26824m	0.416	ppbv	
52) Tetrachloroethene	10.619	164	533289	10.224	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090723\
Data File : VL040680.D
Acq On : 8 Sep 2023 9:51
Operator : SY/MD
Sample : 04323-01DL2 600X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL2

Quant Time: Sep 09 00:09:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Aug 16 05:43:50 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 09/11/2023
Supervised By : Mahesh Dadoda 09/11/2023

