

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090821\
 Data File : VL037572.D
 Acq On : 8 Sep 2021 14:56
 Operator : SY/AP
 Sample : QC090721 CAN 10058
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC090721 CAN 10058

Quant Time: Sep 09 09:12:15 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LTue Aug 17 02:13:47 2021
 QLast Update : Tue Aug 17 02:13:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1615390	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.185	114	4790641	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.098	117	4207269	10.000	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.429	95	3037650	10.163	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090821\
Data File : VL037572.D
Acq On : 8 Sep 2021 14:56
Operator : SY/AP
Sample : QC090721 CAN 10058
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC090721 CAN 10058

Quant Time: Sep 09 09:12:15 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

