

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090821\
 Data File : VL037575.D
 Acq On : 8 Sep 2021 17:00
 Operator : SY/AP
 Sample : M3662-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1DL

Manual Integrations
 APPROVED

MMDadoda
 9/10/2021 10:52:34 AM

Quant Time: Sep 09 09:13:06 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
 QLast Update : Tue Aug 17 02:13:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1658773	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.189	114	4962846	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.098	117	4250874	10.000	ppbv	0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.430	95	2995693	9.919	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.200%
Target Compounds						
					Qvalue	
9) Propene	3.025	41	29364m	0.258	ppbv	
13) Ethanol	3.629	45	118534m	10.718	ppbv	
15) Acetone	3.867	43	237132	1.405	ppbv #	65
19) Isopropyl Alcohol	3.986	45	68390m	0.707	ppbv	
20) Methylene Chloride	4.356	84	46070	0.545	ppbv #	79
26) Hexane	5.697	57	65955	0.315	ppbv #	53
29) Chloroform	5.764	83	88108	0.297	ppbv	95
58) Ethyl Benzene	12.725	91	204301m	0.346	ppbv	
59) m/p-Xylene	12.976	91	1595686m	3.120	ppbv	
60) o-Xylene	13.703	91	660204	1.383	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090821\
Data File : VL037575.D
Acq On : 8 Sep 2021 17:00
Operator : SY/AP
Sample : M3662-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1DL

Manual Integrations
APPROVED

MMDadoda
9/10/2021 10:52:34 AM

Quant Time: Sep 09 09:13:06 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

