

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090821\
 Data File : VL037577.D
 Acq On : 8 Sep 2021 18:25
 Operator : SY/AP
 Sample : M3662-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-1DL

Manual Integrations
 APPROVED

MMDadoda
 9/10/2021 10:52:39 AM

Quant Time: Sep 09 09:13:29 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
 QLast Update : Tue Aug 17 02:13:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1571788	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.185	114	4674828	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.098	117	4092614	10.000	ppbv	0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.426	95	2890236	9.940	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.400%
Target Compounds						
9) Propene	3.025	41	22417m	0.208	ppbv	Qvalue
13) Ethanol	3.619	45	47517m	4.534	ppbv	
15) Acetone	3.867	43	144184	0.901	ppbv #	44
19) Isopropyl Alcohol	3.999	45	29509m	0.322	ppbv	
20) Methylene Chloride	4.349	84	292105	3.647	ppbv	90
26) Hexane	5.697	57	197687	0.996	ppbv #	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090821\
Data File : VL037577.D
Acq On : 8 Sep 2021 18:25
Operator : SY/AP
Sample : M3662-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-1DL

Manual Integrations
APPROVED

MMDadoda
9/10/2021 10:52:39 AM

Quant Time: Sep 09 09:13:29 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

