

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090821\
 Data File : VL037583.D
 Acq On : 8 Sep 2021 22:35
 Operator : SY/AP
 Sample : M3663-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-1DL

Manual Integrations
 APPROVED

MMDadoda
 9/10/2021 10:53:29 AM

Quant Time: Sep 09 09:14:46 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
 QLast Update : Tue Aug 17 02:13:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1547774	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.185	114	4560982	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.098	117	3914666	10.000	ppbv	0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.426	95	2825189	10.158	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.600%
Target Compounds						
					Qvalue	
9) Propene	3.015	41	23845m	0.225	ppbv	
13) Ethanol	3.632	45	39304m	3.809	ppbv	
15) Acetone	3.870	43	180329	1.145	ppbv #	63
20) Methylene Chloride	4.353	84	69843	0.886	ppbv	93
26) Hexane	5.697	57	72796	0.372	ppbv #	47
52) Tetrachloroethene	11.246	164	8125m	0.052	ppbv	
53) Toluene	9.819	91	87543m	0.200	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090821\
Data File : VL037583.D
Acq On : 8 Sep 2021 22:35
Operator : SY/AP
Sample : M3663-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-1DL

Manual Integrations
APPROVED

MMDadoda
9/10/2021 10:53:29 AM

Quant Time: Sep 09 09:14:46 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

