

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091020\
Data File : VL035592.D
Acq On : 10 Sep 2020 20:43
Operator : SY/AP
Sample : CAN 10153
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10153

Quant Time: Sep 11 09:43:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Aug 27 06:46:28 2020
QLast Update : Thu Aug 27 06:46:28 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1020011	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.17	114	4089169	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.09	117	4131930	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	3089801	10.53	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.30%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091020\
Data File : VL035592.D
Acq On : 10 Sep 2020 20:43
Operator : SY/AP
Sample : CAN 10153
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10153

Quant Time: Sep 11 09:43:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Aug 27 06:46:28 2020
QLast Update : Thu Aug 27 06:46:28 2020
Response via : Initial Calibration

