

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041598.D
 Acq On : 16 Sep 2024 17:20
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Sep 17 09:14:19 2024
 Quant Title :
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1175483	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.639	114	3256634	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.449	117	3327610m	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.757 95 2332788 9.110 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 91.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.719	85	1900049	9.396	ppbv	100
3) Chlorodifluoromethane	2.664	51	1860879	13.298	ppbv	99
4) Chloromethane	2.803	50	1044613m	14.035	ppbv	
5) Vinyl Chloride	2.909	62	1021949	13.930	ppbv	99
6) Bromomethane	3.105	94	462431	12.264	ppbv	99
7) Chloroethane	3.185	64	360774	14.249	ppbv	98
8) Dichlorotetrafluoroethane	2.845	85	2287936	12.769	ppbv	99
9) Propene	2.684	41	862437	13.416	ppbv	100
10) Heptane	7.561	43	2181339	15.103	ppbv	100
11) Trichlorofluoromethane	3.549	101	2384942	13.280	ppbv	99
12) 1,1,2-Trichlorotrifluo...	4.050	101	1737916	13.439	ppbv	99
13) Ethanol	3.234	45	32379m	5.636	ppbv	
14) Bromoethene	3.353	108	722975	14.340	ppbv	100
15) Acetone	3.452	43	1746071	13.073	ppbv	99
16) 1,3-Butadiene	2.973	39	997661	13.981	ppbv	99
17) tert-Butyl alcohol	3.861	59	1826662	14.996	ppbv	99
18) 1,1-Dichloroethene	3.864	96	783989	14.172	ppbv	98
19) Isopropyl Alcohol	3.562	45	998621	12.789	ppbv	100
20) Methylene Chloride	3.918	84	642783	12.906	ppbv	93
21) Allyl Chloride	3.979	41	1225956	13.997	ppbv	99
22) trans-1,2-Dichloroethene	4.430	96	781882	14.136	ppbv	99
23) Vinyl Acetate	4.610	43	1212589	11.888	ppbv #	98
24) 1,1-Dichloroethane	4.549	63	1656513	13.792	ppbv	99
25) Ethyl Acetate	5.185	43	3850317	14.511	ppbv	99
26) Hexane	5.195	57	1577948	14.336	ppbv	99
27) Carbon Disulfide	4.108	76	1898447	14.743	ppbv	99
28) Methyl tert-Butyl Ether	4.568	73	934921	13.469	ppbv	100
29) Chloroform	5.259	83	2472134	13.770	ppbv	98
30) Cyclohexane	6.584	84	1384242	15.443	ppbv	96
31) cis-1,2-Dichloroethene	5.063	61	1572482	14.462	ppbv	99
32) 1,1,1-Trichloroethane	5.986	97	2449235	14.015	ppbv	99
34) 2-Butanone	4.767	43	2226315	14.132	ppbv	99
35) Carbon Tetrachloride	6.478	117	2562607	14.267	ppbv	99
36) Benzene	6.352	78	3400759	14.291	ppbv	98
37) 1,2-Dichloroethane	5.790	62	1894411	14.224	ppbv	99
38) Trichloroethene	7.272	130	1777141	16.301	ppbv	97
39) 1,2-Dichloropropane	7.057	63	1318620	14.474	ppbv	99
40) 1,4-Dioxane	7.266	88	453503	12.392	ppbv #	47
41) Tetrahydrofuran	5.536	42	1391996	15.449	ppbv	100
42) Bromodichloromethane	7.233	83	2707351	14.078	ppbv	97
43) Methyl Methacrylate	7.455	69	1368120	16.033	ppbv	98
44) 2,2,4-Trimethylpentane	7.307	57	5619291	14.138	ppbv	99
45) t-1,3-Dichloropropene	8.693	75	1264444	16.099	ppbv	100
46) cis-1,3-Dichloropropene	8.124	75	1733083	15.281	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091624\
 Data File : VL041598.D
 Acq On : 16 Sep 2024 17:20
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Sep 17 09:14:19 2024
 Quant Title :
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
47) 1,1,2-Trichloroethane	8.880	97	1416240	14.103	ppbv	97
48) Dibromochloromethane	9.693	129	2380715	14.716	ppbv	99
49) Bromoform	12.388	173	2103505	15.393	ppbv	100
50) 4-Methyl-2-Pentanone	8.159	43	3768387	15.162	ppbv	99
51) 2-Hexanone	9.529	43	3039864	16.175	ppbv	98
52) Tetrachloroethene	10.597	164	1298509	14.122	ppbv	99
53) Toluene	9.204	91	3987119	15.042	ppbv	99
54) 1,2-Dibromoethane	9.989	107	1951435	14.599	ppbv	100
56) 1,1,1,2-Tetrachloroethane	11.490	131	1724113	13.360	ppbv #	92
57) Chlorobenzene	11.507	112	3100633	13.087	ppbv	98
58) Ethyl Benzene	12.069	91	5323752	13.880	ppbv	100
59) m/p-Xylene	12.352	91	9021854m	27.249	ppbv	
60) o-Xylene	13.037	91	4380559	13.986	ppbv	99
61) Styrene	12.879	104	2207148	15.620	ppbv	100
62) Isopropylbenzene	14.011	105	6290796	13.212	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.027	83	2059305	10.412	ppbv	99
64) n-propylbenzene	14.899	120	1733510	13.926	ppbv	94
65) tert-Butylbenzene	16.072	119	5661694	13.165	ppbv	99
66) Benzyl Chloride	16.304	91	343877	12.466	ppbv	97
67) sec-Butylbenzene	16.616	105	7844186	13.238	ppbv	98
69) p-Isopropyltoluene	16.976	119	6613794	13.451	ppbv	99
70) n-Butylbenzene	17.863	91	6825900	13.542	ppbv	99
71) 2-Chlorotoluene	14.780	91	4774971	13.599	ppbv	99
72) 4-Ethyltoluene	15.172	105	5251979	14.243	ppbv	98
73) 1,3,5-Trimethylbenzene	15.330	105	4527850	14.254	ppbv	98
74) 1,2,4-Trimethylbenzene	16.088	105	4937113	13.368	ppbv	98
75) 1,3-Dichlorobenzene	16.313	146	3113298	13.410	ppbv	99
76) 1,4-Dichlorobenzene	16.455	146	3063921	13.630	ppbv	99
77) 1,2-Dichlorobenzene	17.120	146	3015374	13.102	ppbv	100
78) Hexachloro-1,3-Butadiene	22.609	225	2302119	11.730	ppbv	99
79) Naphthalene	21.358	128	4916804	15.084	ppbv	99
80) Naphthalene,2-methyl-	24.355	142	3899900	14.273	ppbv	99
81) 1,2,4-Trichlorobenzene	21.117	180	2616194	13.487	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

