


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL091719\
Data File : VL034169.D
Acq On : 17 Sep 2019 3:31
Operator : SY/AP
Sample : BFB
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
Last Update : Wed Sep 18 12:53:31 2019



AutoFind: Scans 3642, 3643, 3644; Background Corrected with Scan 3618

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	31.1	259776	PASS
75	95	30	66	63.7	533290	PASS
95	95	100	100	100.0	836608	PASS
96	95	5	9	7.0	58954	PASS
173	174	0.00	2	0.9	3987	PASS
174	95	50	120	55.9	467840	PASS
175	174	4	9	6.8	31874	PASS
176	174	93	101	96.8	452906	PASS
177	176	5	9	7.1	32341	PASS