

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091819\
 Data File : VL034187.D
 Acq On : 18 Sep 2019 17:48
 Operator : SY/AP
 Sample : OC CAN 10281
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC CAN 10281

Manual Integrations
 APPROVED

MMDadoda
 9/20/2019 12:37:17 AM

Quant Time: Sep 19 06:59:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Sep 18 12:53:31 2019
 QLast Update : Wed Sep 18 12:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.70	49	1200240	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.22	114	1968173	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.16	117	1980151	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.50	95	1776006	10.32	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	7220m	0.429	ppbv	
15) Acetone	3.86	43	47429m	0.295	ppbv	
20) Methylene Chloride	4.35	84	28131	0.607	ppbv #	85
26) Hexane	5.72	57	18977m	0.150	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091819\
Data File : VL034187.D
Acq On : 18 Sep 2019 17:48
Operator : SY/AP
Sample : OC CAN 10281
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC CAN 10281

Manual Integrations
APPROVED

MMDadoda
9/20/2019 12:37:17 AM

Quant Time: Sep 19 06:59:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL091719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep
QLast Update : Wed Sep 18 12:53:31 2019
Response via : Initial Calibration

