

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
 Data File : VL041615.D
 Acq On : 18 Sep 2024 20:38
 Operator : SY/MD
 Sample : QC090524 CAN10153
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC090524 CAN10153

Quant Time: Sep 19 03:40:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/19/2024
 Supervised By :Mahesh Dadoda 09/19/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	1000091	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	6.655	114	2767195	10.000	ppbv	0.03
55) Chlorobenzene-d5	11.465	117	2435639	10.000	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.773	95	1757273	9.801	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091824\
Data File : VL041615.D
Acq On : 18 Sep 2024 20:38
Operator : SY/MD
Sample : QC090524 CAN10153
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC090524 CAN10153

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/19/2024
Supervised By :Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 03:40:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration

