

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092222\
 Data File : VL039671.D
 Acq On : 22 Sep 2022 18:19
 Operator : SY/AP
 Sample : N4669-14DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0PG6DL

Quant Time: Sep 23 06:36:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Sep 19 14:26:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.224	49	991022	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.684	114	2587406	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.497	117	2319521	10.000	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.802	95	1714967	9.904	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.000%
Target Compounds						
					Qvalue	
13) Ethanol	3.285	45	18237	4.456	ppbv	82
15) Acetone	3.517	43	139663	1.266	ppbv	95
20) Methylene Chloride	3.973	84	50945	0.999	ppbv	91
38) Trichloroethene	7.327	130	15703	0.147	ppbv #	86
52) Tetrachloroethene	10.645	164	269940	3.577	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092222\
Data File : VL039671.D
Acq On : 22 Sep 2022 18:19
Operator : SY/AP
Sample : N4669-14DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0PG6DL

Quant Time: Sep 23 06:36:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Sep 19 14:26:28 2022
Response via : Initial Calibration

