

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092222\
 Data File : VL039674.D
 Acq On : 22 Sep 2022 20:14
 Operator : SY/AP
 Sample : N4669-12DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0PG4DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/27/2022
 Supervised By :Mahesh Dadoda 09/28/2022

Quant Time: Sep 23 06:37:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Sep 19 14:26:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.221	49	941074	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.677	114	2424618	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.491	117	2199218	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.799	95	1644253	10.015	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.200%
Target Compounds						
					Qvalue	
13) Ethanol	3.272	45	222646m	57.290	ppbv	
15) Acetone	3.504	43	231587	2.212	ppbv	97
20) Methylene Chloride	3.967	84	237105m	4.897	ppbv	
26) Hexane	5.246	57	89675	0.925	ppbv #	43
38) Trichloroethene	7.317	130	21068m	0.210	ppbv	
52) Tetrachloroethene	10.638	164	331924	4.693	ppbv	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092222\
Data File : VL039674.D
Acq On : 22 Sep 2022 20:14
Operator : SY/AP
Sample : N4669-12DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0PG4DL

Quant Time: Sep 23 06:37:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Sep 19 14:26:28 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/27/2022
Supervised By :Mahesh Dadoda 09/28/2022

