

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL092320\
 Data File : VL035709.D
 Acq On : 23 Sep 2020 6:30
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

MMDadoda
 9/24/2020 1:35:21 PM

Quant Time: Sep 23 07:07:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL092320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Sep 23 07:07:44 2020
 QLast Update : Wed Sep 23 05:25:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1183930	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	4309510	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	4341510	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	3255169	10.43	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.25	62	2478m	0.029	ppbv	
32) 1,1,1-Trichloroethane	6.52	97	5756m	0.023	ppbv	
35) Carbon Tetrachloride	7.03	117	6136m	0.028	ppbv	
38) Trichloroethene	7.85	130	4037m	0.025	ppbv	
52) Tetrachloroethene	11.25	164	4205m	0.026	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.70	83	10367m	0.035	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL092320\
Data File : VL035709.D
Acq On : 23 Sep 2020 6:30
Operator : SY/AP
Sample : VSTDIC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03

Manual Integrations
APPROVED

MMDadoda
9/24/2020 1:35:21 PM

Quant Time: Sep 23 07:07:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL092320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep
QLast Update : Wed Sep 23 05:25:46 2020
Response via : Initial Calibration

