

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL092320\
 Data File : VL035715.D
 Acq On : 23 Sep 2020 12:23
 Operator : SY/AP
 Sample : L4079-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 B-1DL

Quant Time: Sep 24 11:29:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL092320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 23 07:38:05 2020
 QLast Update : Wed Sep 23 07:38:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1271072	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	4657227	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	4650627	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	3510774	10.07	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	21256	1.736	ppbv	97
26) Hexane	5.70	57	67668	0.502	ppbv #	50
44) 2,2,4-Trimethylpentane	7.89	57	355689	0.817	ppbv	95
53) Toluene	9.82	91	37443	0.105	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL092320\
Data File : VL035715.D
Acq On : 23 Sep 2020 12:23
Operator : SY/AP
Sample : L4079-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
B-1DL

Quant Time: Sep 24 11:29:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL092320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 23 07:38:05 2020
QLast Update : Wed Sep 23 07:38:05 2020
Response via : Initial Calibration

