

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL092320\
 Data File : VL035718.D
 Acq On : 23 Sep 2020 14:24
 Operator : SY/AP
 Sample : L4079-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 B-2DL

Manual Integrations
 APPROVED

MMDadoda
 9/24/2020 1:35:38 PM

Quant Time: Sep 24 08:59:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL092320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 23 2020
 QLast Update : Wed Sep 23 07:38:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1260034	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	4752992	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	4780861	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.45	95	3560194	9.94	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	12356	0.251	ppbv #	80
10) Heptane	8.14	43	13924	0.101	ppbv #	19
13) Ethanol	3.62	45	17324	1.427	ppbv	94
15) Acetone	3.86	43	50085m	0.369	ppbv	
19) Isopropyl Alcohol	3.99	45	21384	0.290	ppbv #	89
20) Methylene Chloride	4.35	84	94884	0.953	ppbv	97
26) Hexane	5.70	57	70290	0.526	ppbv #	51
44) 2,2,4-Trimethylpentane	7.89	57	433158	0.975	ppbv	96
52) Tetrachloroethene	11.24	164	5404m	0.036	ppbv	
53) Toluene	9.83	91	36059	0.099	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL092320\
Data File : VL035718.D
Acq On : 23 Sep 2020 14:24
Operator : SY/AP
Sample : L4079-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
B-2DL

Manual Integrations
APPROVED

MMDadoda
9/24/2020 1:35:38 PM

Quant Time: Sep 24 08:59:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL092320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep
QLast Update : Wed Sep 23 07:38:05 2020
Response via : Initial Calibration

