

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/24/2025
 Supervised By : Mahesh Dadoda 09/26/2025

Quant Time: Sep 24 02:04:07 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.793	49	140159	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	397606	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.888	117	356928	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.383	95	285330	10.709	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	= 107.100%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	187032	9.921	ppbv	95
3) Chlorodifluoromethane	1.479	51	226561	9.643	ppbv	97
4) Chloromethane	1.534	50	78049	9.147	ppbv	99
5) Vinyl Chloride	1.583	62	67278	8.722	ppbv	100
6) Bromomethane	1.670	94	24422	8.975	ppbv	96
7) Chloroethane	1.709	64	26802	8.948	ppbv	98
8) Dichlorotetrafluoroethane	1.557	85	151817	9.434	ppbv	100
9) Propene	1.486	41	83019	8.122	ppbv	96
10) Heptane	4.991	43	242117	8.876	ppbv	99
11) Trichlorofluoromethane	1.881	101	185063	10.181	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.143	101	144385	9.809	ppbv	99
13) Ethanol	1.725	45	6504	6.559	ppbv	96
14) Bromoethene	1.787	108	51728	9.295	ppbv	98
15) Acetone	1.839	43	146605	8.031	ppbv	98
16) 1,3-Butadiene	1.612	39	74090	8.604	ppbv	97
17) tert-Butyl alcohol	2.043	59	167840	8.383	ppbv	100
18) 1,1-Dichloroethene	2.039	96	63829	9.368	ppbv	99
19) Isopropyl Alcohol	1.890	45	94735	8.633	ppbv	95
20) Methylene Chloride	2.065	84	55766	10.264	ppbv	96
21) Allyl Chloride	2.101	41	117413	9.235	ppbv	97
22) trans-1,2-Dichloroethene	2.344	96	71690	9.260	ppbv	97
23) Vinyl Acetate	2.467	43	230988	8.857	ppbv #	98
24) 1,1-Dichloroethane	2.415	63	145685	9.596	ppbv	99
25) Ethyl Acetate	2.839	43	423103	9.170	ppbv	99
26) Hexane	2.832	57	188410	8.959	ppbv	98
27) Carbon Disulfide	2.156	76	170121	9.019	ppbv	96
28) Methyl tert-Butyl Ether	2.444	73	91757	8.393	ppbv	95
29) Chloroform	2.855	83	281560	10.239	ppbv	99
30) Cyclohexane	3.862	84	158650	9.022	ppbv	97
31) cis-1,2-Dichloroethene	2.725	61	188239	9.555	ppbv	100
32) 1,1,1-Trichloroethane	3.373	97	274087	9.228	ppbv	99
34) 2-Butanone	2.560	43	279284	10.023	ppbv	99
35) Carbon Tetrachloride	3.771	117	279346	10.791	ppbv	98
36) Benzene	3.664	78	373905	10.014	ppbv	99
37) 1,2-Dichloroethane	3.224	62	224338	11.424	ppbv	97
38) Trichloroethene	4.557	130	174937	9.427	ppbv	92
39) 1,2-Dichloropropane	4.324	63	138847	10.186	ppbv	87
40) 1,4-Dioxane	4.587	88	58847	9.285	ppbv #	97
41) Tetrahydrofuran	3.059	42	152129	9.498	ppbv	100
42) Bromodichloromethane	4.499	83	297856	10.891	ppbv	97
43) Methyl Methacrylate	4.891	69	143508	9.394	ppbv	99
44) 2,2,4-Trimethylpentane	4.648	57	637574	9.887	ppbv	97
45) t-1,3-Dichloropropene	6.428	75	149384	9.169	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/24/2025
 Supervised By :Mahesh Dadoda 09/26/2025

Quant Time: Sep 24 02:04:07 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 23 02:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.613	75	197209	9.498	ppbv	96
47) 1,1,2-Trichloroethane	6.590	97	146175	10.149	ppbv	97
48) Dibromochloromethane	7.396	129	241948	10.500	ppbv	99
49) Bromoform	9.490	173	201373	10.047	ppbv	100
50) 4-Methyl-2-Pentanone	5.758	43	371641	9.951	ppbv	99
51) 2-Hexanone	7.444	43	299396	10.057	ppbv #	97
52) Tetrachloroethene	8.231	164	130498	9.033	ppbv	97
53) Toluene	6.943	91	424755	9.612	ppbv	100
54) 1,2-Dibromoethane	7.661	107	210744	10.396	ppbv	96
56) 1,1,1,2-Tetrachloroethane	8.933	131	181159	10.604	ppbv	98
57) Chlorobenzene	8.930	112	332694	10.210	ppbv	99
58) Ethyl Benzene	9.364	91	602544	10.281	ppbv	98
59) m/p-Xylene	9.558	91	964478m	20.873	ppbv	
60) o-Xylene	9.972	91	484224	10.485	ppbv	99
61) Styrene	9.878	104	212504	9.728	ppbv	97
62) Isopropylbenzene	10.545	105	703444	10.219	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.969	83	246998	10.175	ppbv	100
64) n-propylbenzene	10.995	120	181920	10.113	ppbv	94
65) tert-Butylbenzene	11.536	119	612510	10.004	ppbv	97
66) Benzyl Chloride	11.623	91	38429	5.982	ppbv	98
67) sec-Butylbenzene	11.772	105	877873	10.290	ppbv	98
69) p-Isopropyltoluene	11.930	119	733573	10.179	ppbv	98
70) n-Butylbenzene	12.286	91	744721	10.584	ppbv	99
71) 2-Chlorotoluene	10.908	91	531933	10.234	ppbv	98
72) 4-Ethyltoluene	11.131	105	592733	10.248	ppbv	98
73) 1,3,5-Trimethylbenzene	11.209	105	491234	10.472	ppbv	97
74) 1,2,4-Trimethylbenzene	11.542	105	545008	10.521	ppbv	96
75) 1,3-Dichlorobenzene	11.613	146	327787	10.203	ppbv	99
76) 1,4-Dichlorobenzene	11.678	146	328439	10.293	ppbv	98
77) 1,2-Dichlorobenzene	11.953	146	313228	10.247	ppbv	99
78) Hexachloro-1,3-Butadiene	13.885	225	214126	8.956	ppbv	99
79) Naphthalene	13.516	128	503948	11.001	ppbv	99
80) Naphthalene,2-methyl-	14.462	142	125827	8.892	ppbv	95
81) 1,2,4-Trichlorobenzene	13.445	180	261685	10.684	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Quant Time: Sep 24 02:04:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Tue Sep 23 02:39:16 2025
Response via : Initial Calibration

