

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 24 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	1.345	1.334	0.8	88	0.00
3	Chlorodifluoromethane	1.676	1.616	3.6	84	0.00
4	Chloromethane	0.609	0.557	8.5	82	0.00
5 T	Vinyl Chloride	0.550	0.480	12.7	84	0.00
6 T	Bromomethane	0.194	0.174	10.3	86	0.00
7	Chloroethane	0.214	0.191	10.7	80	0.00
8 T	Dichlorotetrafluoroethane	1.148	1.083	5.7	85	0.00
9 T	Propene	0.729	0.592	18.8	71	0.00
10 T	Heptane	1.946	1.727	11.3	76	0.02
11 T	Trichlorofluoromethane	1.297	1.320	-1.8	90	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.050	1.030	1.9	88	0.00
13	Ethanol	0.071	0.046#	35.2#	72	0.00
14 T	Bromoethene	0.397	0.369	7.1	84	0.00
15 T	Acetone	1.302	1.046	19.7	86	0.00
16 T	1,3-Butadiene	0.614	0.529	13.8	83	0.00
17	tert-Butyl alcohol	1.429	1.197	16.2	73	0.00
18 T	1,1-Dichloroethene	0.486	0.455	6.4	84	0.00
19 T	Isopropyl Alcohol	0.783	0.676	13.7	77	0.00
20 T	Methylene Chloride	0.545	0.398	27.0	84	0.00
21 T	Allyl Chloride	0.907	0.838	7.6	80	0.00
22 T	trans-1,2-Dichloroethene	0.552	0.511	7.4	83	0.00
23 T	Vinyl Acetate	1.861	1.648	11.4	85	0.00
24 T	1,1-Dichloroethane	1.083	1.039	4.1	86	0.00
25 T	Ethyl Acetate	3.292	3.019	8.3	79	0.00
26 T	Hexane	1.501	1.344	10.5	79	0.00
27 T	Carbon Disulfide	1.346	1.214	9.8	78	0.00
28 T	Methyl tert-Butyl Ether	0.780	0.655	16.0	84	0.00
29 T	Chloroform	1.962	2.009	-2.4	89	0.00
30 T	Cyclohexane	1.255	1.132	9.8	78	0.00
31 T	cis-1,2-Dichloroethene	1.406	1.343	4.5	81	0.00
32 T	1,1,1-Trichloroethane	2.119	1.956	7.7	84	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.01
34 T	2-Butanone	0.701	0.702	-0.1	81	0.00
35 T	Carbon Tetrachloride	0.651	0.703	-8.0	87	0.02
36 T	Benzene	0.939	0.940	-0.1	79	0.01
37 T	1,2-Dichloroethane	0.494	0.564	-14.2	91	0.00
38 T	Trichloroethene	0.467	0.440	5.8	78	0.02
39 T	1,2-Dichloropropane	0.343	0.349	-1.7	81	0.02
40 T	1,4-Dioxane	0.159	0.148	6.9	76	0.02
41 T	Tetrahydrofuran	0.403	0.383	5.0	74	0.01
42 T	Bromodichloromethane	0.688	0.749	-8.9	83	0.02
43	Methyl Methacrylate	0.384	0.361	6.0	74	0.02
44 T	2,2,4-Trimethylpentane	1.622	1.604	1.1	77	0.02
45 T	t-1,3-Dichloropropene	0.410	0.376	8.3	69	0.02
46 T	cis-1,3-Dichloropropene	0.522	0.496	5.0	72	0.02
47 T	1,1,2-Trichloroethane	0.362	0.368	-1.7	81	0.02
48 T	Dibromochloromethane	0.580	0.609	-5.0	81	0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 24 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.504	0.506	-0.4	75	0.00
50 T	4-Methyl-2-Pentanone	0.939	0.935	0.4	76	0.02
51 T	2-Hexanone	0.749	0.753	-0.5	76	0.02
52 T	Tetrachloroethene	0.363	0.328	9.6	76	0.01
53 T	Toluene	1.111	1.068	3.9	75	0.02
54 T	1,2-Dibromoethane	0.510	0.530	-3.9	82	0.02
55 I	Chlorobenzene-d5	1.000	1.000	0.0	73	0.01
56	1,1,1,2-Tetrachloroethane	0.479	0.508	-6.1	81	0.01
57 T	Chlorobenzene	0.913	0.932	-2.1	80	0.01
58 T	Ethyl Benzene	1.642	1.688	-2.8	79	0.02
59 T	m/p-Xylene	1.295	1.351	-4.3	81	0.03
60 T	o-Xylene	1.294	1.357	-4.9	81	0.01
61 T	Styrene	0.612	0.595	2.8	72	0.01
62	Isopropylbenzene	1.929	1.971	-2.2	80	0.00
63 T	1,1,2,2-Tetrachloroethane	0.680	0.692	-1.8	83	0.01
64	n-propylbenzene	0.504	0.510	-1.2	79	0.00
65	tert-Butylbenzene	1.715	1.716	-0.1	81	0.00
66 T	Benzyl Chloride	0.180	0.108	40.0#	40	0.01
67	sec-Butylbenzene	2.390	2.460	-2.9	82	0.00
68 S	1-Bromo-4-Fluorobenzene	0.746	0.799	-7.1	77	0.01
69	p-Isopropyltoluene	2.019	2.055	-1.8	81	0.00
70	n-Butylbenzene	1.971	2.086	-5.8	81	0.00
71	2-Chlorotoluene	1.456	1.490	-2.3	80	0.01
72 T	4-Ethyltoluene	1.621	1.661	-2.5	80	0.00
73 T	1,3,5-Trimethylbenzene	1.314	1.376	-4.7	80	0.00
74 T	1,2,4-Trimethylbenzene	1.451	1.527	-5.2	83	0.00
75 T	1,3-Dichlorobenzene	0.900	0.918	-2.0	80	0.00
76 T	1,4-Dichlorobenzene	0.894	0.920	-2.9	80	0.00
77 T	1,2-Dichlorobenzene	0.856	0.878	-2.6	81	0.00
78 T	Hexachloro-1,3-Butadiene	0.670	0.600	10.4	76	0.00
79 T	Naphthalene	1.283	1.412	-10.1	75	0.00
80 T	Naphthalene,2-methyl-	0.396	0.353	10.9	64	0.00
81 T	1,2,4-Trichlorobenzene	0.686	0.733	-6.9	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0