

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 24 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	10.000	9.921	0.8	88	0.00
3	Chlorodifluoromethane	10.000	9.643	3.6	84	0.00
4	Chloromethane	10.000	9.147	8.5	82	0.00
5 T	Vinyl Chloride	10.000	8.722	12.8	84	0.00
6 T	Bromomethane	10.000	8.975	10.3	86	0.00
7	Chloroethane	10.000	8.948	10.5	80	0.00
8 T	Dichlorotetrafluoroethane	10.000	9.434	5.7	85	0.00
9 T	Propene	10.000	8.122	18.8	71	0.00
10 T	Heptane	10.000	8.876	11.2	76	0.02
11 T	Trichlorofluoromethane	10.000	10.181	-1.8	90	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	9.809	1.9	88	0.00
13	Ethanol	10.000	6.559	34.4#	72	0.00
14 T	Bromoethene	10.000	9.295	7.1	84	0.00
15 T	Acetone	10.000	8.031	19.7	86	0.00
16 T	1,3-Butadiene	10.000	8.604	14.0	83	0.00
17	tert-Butyl alcohol	10.000	8.383	16.2	73	0.00
18 T	1,1-Dichloroethene	10.000	9.368	6.3	84	0.00
19 T	Isopropyl Alcohol	10.000	8.633	13.7	77	0.00
20 T	Methylene Chloride	10.000	10.264	-2.6	84	0.00
21 T	Allyl Chloride	10.000	9.235	7.7	80	0.00
22 T	trans-1,2-Dichloroethene	10.000	9.260	7.4	83	0.00
23 T	Vinyl Acetate	10.000	8.857	11.4	85	0.00
24 T	1,1-Dichloroethane	10.000	9.596	4.0	86	0.00
25 T	Ethyl Acetate	10.000	9.170	8.3	79	0.00
26 T	Hexane	10.000	8.959	10.4	79	0.00
27 T	Carbon Disulfide	10.000	9.019	9.8	78	0.00
28 T	Methyl tert-Butyl Ether	10.000	8.393	16.1	84	0.00
29 T	Chloroform	10.000	10.239	-2.4	89	0.00
30 T	Cyclohexane	10.000	9.022	9.8	78	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.555	4.5	81	0.00
32 T	1,1,1-Trichloroethane	10.000	9.228	7.7	84	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	75	0.01
34 T	2-Butanone	10.000	10.023	-0.2	81	0.00
35 T	Carbon Tetrachloride	10.000	10.791	-7.9	87	0.02
36 T	Benzene	10.000	10.014	-0.1	79	0.01
37 T	1,2-Dichloroethane	10.000	11.424	-14.2	91	0.00
38 T	Trichloroethene	10.000	9.427	5.7	78	0.02
39 T	1,2-Dichloropropane	10.000	10.186	-1.9	81	0.02
40 T	1,4-Dioxane	10.000	9.285	7.1	76	0.02
41 T	Tetrahydrofuran	10.000	9.498	5.0	74	0.01
42 T	Bromodichloromethane	10.000	10.891	-8.9	83	0.02
43	Methyl Methacrylate	10.000	9.394	6.1	74	0.02
44 T	2,2,4-Trimethylpentane	10.000	9.887	1.1	77	0.02
45 T	t-1,3-Dichloropropene	10.000	9.169	8.3	69	0.02
46 T	cis-1,3-Dichloropropene	10.000	9.498	5.0	72	0.02
47 T	1,1,2-Trichloroethane	10.000	10.149	-1.5	81	0.02
48 T	Dibromochloromethane	10.000	10.500	-5.0	81	0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 24 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	10.047	-0.5	75	0.00
50 T	4-Methyl-2-Pentanone	10.000	9.951	0.5	76	0.02
51 T	2-Hexanone	10.000	10.057	-0.6	76	0.02
52 T	Tetrachloroethene	10.000	9.033	9.7	76	0.01
53 T	Toluene	10.000	9.612	3.9	75	0.02
54 T	1,2-Dibromoethane	10.000	10.396	-4.0	82	0.02
55 I	Chlorobenzene-d5	10.000	10.000	0.0	73	0.01
56	1,1,1,2-Tetrachloroethane	10.000	10.604	-6.0	81	0.01
57 T	Chlorobenzene	10.000	10.210	-2.1	80	0.01
58 T	Ethyl Benzene	10.000	10.281	-2.8	79	0.02
59 T	m/p-Xylene	20.000	20.873	-4.4	81	0.03
60 T	o-Xylene	10.000	10.485	-4.8	81	0.01
61 T	Styrene	10.000	9.728	2.7	72	0.01
62	Isopropylbenzene	10.000	10.219	-2.2	80	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	10.175	-1.8	83	0.01
64	n-propylbenzene	10.000	10.113	-1.1	79	0.00
65	tert-Butylbenzene	10.000	10.004	-0.0	81	0.00
66 T	Benzyl Chloride	10.000	5.982	40.2#	40	0.01
67	sec-Butylbenzene	10.000	10.290	-2.9	82	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.709	-7.1	77	0.01
69	p-Isopropyltoluene	10.000	10.179	-1.8	81	0.00
70	n-Butylbenzene	10.000	10.584	-5.8	81	0.00
71	2-Chlorotoluene	10.000	10.234	-2.3	80	0.01
72 T	4-Ethyltoluene	10.000	10.248	-2.5	80	0.00
73 T	1,3,5-Trimethylbenzene	10.000	10.472	-4.7	80	0.00
74 T	1,2,4-Trimethylbenzene	10.000	10.521	-5.2	83	0.00
75 T	1,3-Dichlorobenzene	10.000	10.203	-2.0	80	0.00
76 T	1,4-Dichlorobenzene	10.000	10.293	-2.9	80	0.00
77 T	1,2-Dichlorobenzene	10.000	10.247	-2.5	81	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	8.956	10.4	76	0.00
79 T	Naphthalene	10.000	11.001	-10.0	75	0.00
80 T	Naphthalene,2-methyl-	10.000	8.892	11.1	64	0.00
81 T	1,2,4-Trichlorobenzene	10.000	10.684	-6.8	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0