

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL092419\
 Data File : VL034203.D
 Acq On : 24 Sep 2019 16:34
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 25 06:59:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
 QLast Update : Wed Sep 18 12:53:31 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	1.477	1.148	22.3	82	0.00
3	Chlorodifluoromethane	1.173	1.093	6.8	88	0.00
4	Chloromethane	0.689	0.674	2.2	91	0.00
5 T	Vinyl Chloride	0.630	0.599	4.9	91	0.00
6 T	Bromomethane	0.294	0.297	-1.0	92	0.00
7	Chloroethane	0.220	0.215	2.3	90	0.00
8 T	Dichlorotetrafluoroethane	1.304	1.246	4.4	93	0.00
9 T	Propene	0.578	0.566	2.1	90	0.00
10 T	Heptane	1.508	1.454	3.6	88	0.00
11 T	Trichlorofluoromethane	1.345	1.302	3.2	92	0.00
12 T	1,1,2-Trichlorotrifluoroeth	0.880	0.865	1.7	93	0.00
13	Ethanol	0.140	0.124	11.4	120	0.00
14 T	Bromoethene	0.360	0.371	-3.1	92	0.00
15 T	Acetone	1.340	1.111	17.1	92	0.00
16 T	1,3-Butadiene	0.585	0.582	0.5	90	0.00
17	tert-Butyl alcohol	0.849	1.013	-19.3	110	0.00
18 T	1,1-Dichloroethene	0.384	0.393	-2.3	92	0.00
19 T	Isopropyl Alcohol	0.824	0.857	-4.0	109	0.00
20 T	Methylene Chloride	0.386	0.371	3.9	97	0.00
21 T	Allyl Chloride	0.757	0.767	-1.3	89	0.00
22 T	trans-1,2-Dichloroethene	0.384	0.392	-2.1	92	0.00
23 T	Vinyl Acetate	1.975	1.923	2.6	89	0.00
24 T	1,1-Dichloroethane	1.330	1.251	5.9	89	0.00
25 T	Ethyl Acetate	2.534	2.350	7.3	91	0.00
26 T	Hexane	1.054	1.009	4.3	92	0.00
27 T	Carbon Disulfide	0.972	1.054	-8.4	91	0.00
28 T	Methyl tert-Butyl Ether	1.566	1.613	-3.0	95	0.00
29 T	Chloroform	1.507	1.459	3.2	90	0.00
30 T	Cyclohexane	0.761	0.760	0.1	90	0.00
31 T	cis-1,2-Dichloroethene	1.075	1.066	0.8	89	0.00
32 T	1,1,1-Trichloroethane	1.453	1.384	4.7	89	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
34 T	2-Butanone	1.021	0.915	10.4	87	0.00
35 T	Carbon Tetrachloride	0.831	0.770	7.3	89	0.00
36 T	Benzene	1.104	1.039	5.9	90	0.00
37 T	1,2-Dichloroethane	0.786	0.720	8.4	86	0.00
38 T	Trichloroethene	0.395	0.349	11.6	89	0.00
39 T	1,2-Dichloropropane	0.496	0.466	6.0	90	0.00
40 T	1,4-Dioxane	0.161	0.156	3.1	104	0.00
41 T	Tetrahydrofuran	0.575	0.560	2.6	89	0.00
42 T	Bromodichloromethane	0.943	0.903	4.2	88	0.00
43	Methyl Methacrylate	0.456	0.453	0.7	89	0.00
44 T	2,2,4-Trimethylpentane	2.126	1.924	9.5	91	0.00
45 T	t-1,3-Dichloropropene	0.539	0.590	-9.5	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL092419\
 Data File : VL034203.D
 Acq On : 24 Sep 2019 16:34
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 25 06:59:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
 QLast Update : Wed Sep 18 12:53:31 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.649	0.666	-2.6	87	0.00
47 T	1,1,2-Trichloroethane	0.446	0.412	7.6	89	0.00
48 T	Dibromochloromethane	0.706	0.684	3.1	88	0.00
49 T	Bromoform	0.572	0.563	1.6	86	0.00
50 T	4-Methyl-2-Pentanone	1.486	1.349	9.2	88	0.00
51 T	2-Hexanone	1.247	1.149	7.9	86	0.00
52 T	Tetrachloroethene	0.365	0.325	11.0	90	0.00
53 T	Toluene	1.207	1.159	4.0	90	0.00
54 T	1,2-Dibromoethane	0.663	0.622	6.2	89	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
56	1,1,1,2-Tetrachloroethane	0.446	0.419	6.1	88	0.00
57 T	Chlorobenzene	0.871	0.762	12.5	88	0.00
58 T	Ethyl Benzene	1.614	1.472	8.8	89	0.00
59 T	m/p-Xylene	1.336	1.189	11.0	89	0.00
60 T	o-Xylene	1.336	1.185	11.3	89	0.00
61 T	Styrene	0.957	0.897	6.3	88	0.00
62	Isopropylbenzene	1.730	1.529	11.6	89	0.00
63 T	1,1,2,2-Tetrachloroethane	1.096	0.867	20.9	89	0.00
64	n-propylbenzene	0.438	0.389	11.2	85	0.00
65	tert-Butylbenzene	1.512	1.294	14.4	88	0.00
66 T	Benzyl Chloride	0.535	0.615	-15.0	84	0.00
67	sec-Butylbenzene	2.161	1.878	13.1	89	0.00
68 S	1-Bromo-4-Fluorobenzene	0.869	0.854	1.7	86	0.00
69	p-Isopropyltoluene	1.750	1.519	13.2	88	0.00
70	n-Butylbenzene	2.038	1.727	15.3	88	0.00
71	2-Chlorotoluene	1.413	1.258	11.0	88	0.00
72 T	4-Ethyltoluene	1.510	1.361	9.9	88	0.00
73 T	1,3,5-Trimethylbenzene	1.459	1.297	11.1	88	0.00
74 T	1,2,4-Trimethylbenzene	1.549	1.306	15.7	89	0.00
75 T	1,3-Dichlorobenzene	0.831	0.702	15.5	87	0.00
76 T	1,4-Dichlorobenzene	0.851	0.709	16.7	85	0.00
77 T	1,2-Dichlorobenzene	0.841	0.692	17.7	85	0.00
78 T	Hexachloro-1,3-Butadiene	0.557	0.527	5.4	105	0.00
79 T	Naphthalene	1.363	1.172	14.0	87	0.00
80 T	Naphthalene,2-methyl-	0.634	0.518	18.3	70	0.00
81 T	1,2,4-Trichlorobenzene	0.755	0.635	15.9	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0