


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL092920\
 Data File : VL035723.D
 Acq On : 29 Sep 2020 9:01
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL092320AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Sep 23 07:38:05 2020
 Last Update : Wed Sep 23 07:38:05 2020



AutoFind: Scans 3624, 3625, 3626; Background Corrected with Scan 3600

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	16.1	251242	PASS
75	95	30	66	48.9	764821	PASS
95	95	100	100	100.0	1564160	PASS
96	95	5	9	7.4	115061	PASS
173	174	0.00	2	0.5	6194	PASS
174	95	50	120	76.5	1196373	PASS
175	174	4	9	7.4	88866	PASS
176	174	93	101	96.3	1151829	PASS
177	176	5	9	7.1	81672	PASS