

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL100121\
 Data File : VL037746.D
 Acq On : 1 Oct 2021 18:02
 Operator : SY/AP
 Sample : M4012-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P17DL

Manual Integrations
 APPROVED

MMDadoda
 10/4/2021 5:50:05 PM

Quant Time: Oct 02 19:46:46 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 1
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1463468	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4323021	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3729693	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.40	95	2489142	9.23	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	132122m	13.872	ppbv	
15) Acetone	3.84	43	206184	1.356	ppbv	# 86
19) Isopropyl Alcohol	3.96	45	497270	6.011	ppbv	98
20) Methylene Chloride	4.33	84	128907	1.776	ppbv	94
29) Chloroform	5.74	83	561957	2.200	ppbv	95
52) Tetrachloroethene	11.20	164	330588	2.389	ppbv	97
75) 1,3-Dichlorobenzene	16.98	146	37480m	0.130	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL100121\
Data File : VL037746.D
Acq On : 1 Oct 2021 18:02
Operator : SY/AP
Sample : M4012-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P17DL

Manual Integrations
APPROVED

MMDadoda
10/4/2021 5:50:05 PM

Quant Time: Oct 02 19:46:46 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

