

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL100121\
 Data File : VL037749.D
 Acq On : 1 Oct 2021 19:57
 Operator : SY/AP
 Sample : M4012-06DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P28DL

Manual Integrations
 APPROVED

MMDadoda
 10/4/2021 5:50:10 PM

Quant Time: Oct 02 19:50:30 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Sep 1
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1465604	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4242843	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	3508651	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2393387	9.44	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.63	45	38128m	3.997	ppbv	
15) Acetone	3.85	43	183491m	1.205	ppbv	
20) Methylene Chloride	4.34	84	71687	0.986	ppbv	87
59) m/p-Xylene	12.96	91	40649m	0.100	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL100121\
Data File : VL037749.D
Acq On : 1 Oct 2021 19:57
Operator : SY/AP
Sample : M4012-06DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P28DL

Manual Integrations
APPROVED

MMDadoda
10/4/2021 5:50:10 PM

Quant Time: Oct 02 19:50:30 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

