

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL100121\
 Data File : VL037751.D
 Acq On : 1 Oct 2021 21:13
 Operator : SY/AP
 Sample : M4012-07DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P09DL

Manual Integrations
 APPROVED

apatel
 10/14/2021 6:07:32 AM

Quant Time: Oct 02 19:52:15 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 1
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1398282	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4078852	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	3388834	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.39	95	2395772	9.78	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.84	43	169351	1.166	ppbv	# 77
20) Methylene Chloride	4.34	84	37621	0.542	ppbv	91
52) Tetrachloroethene	11.20	164	442650	3.390	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL100121\
Data File : VL037751.D
Acq On : 1 Oct 2021 21:13
Operator : SY/AP
Sample : M4012-07DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P09DL

Manual Integrations
APPROVED

apatel
10/14/2021 6:07:32 AM

Quant Time: Oct 02 19:52:15 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

