


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL100219\
Data File : VL034204.D
Acq On : 2 Oct 2019 9:01
Operator : SY/AP
Sample : BFB
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
Last Update : Wed Sep 18 12:53:31 2019



AutoFind: Scans 3639, 3640, 3641; Background Corrected with Scan 3615

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	30.7	225216	PASS
75	95	30	66	62.9	460906	PASS
95	95	100	100	100.0	732608	PASS
96	95	5	9	7.3	53442	PASS
173	174	0.00	2	0.5	2036	PASS
174	95	50	120	58.2	426218	PASS
175	174	4	9	7.7	32856	PASS
176	174	93	101	95.6	407360	PASS
177	176	5	9	7.3	29813	PASS