

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100422\
 Data File : VL039682.D
 Acq On : 4 Oct 2022 15:51
 Operator : SY/AP
 Sample : VL1004ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1004ABL01

Quant Time: Oct 05 04:39:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Sep 19 14:26:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	1093043	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.668	114	2845634	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.474	117	2575439	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.783	95	1862668	9.688	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100422\
Data File : VL039682.D
Acq On : 4 Oct 2022 15:51
Operator : SY/AP
Sample : VL1004ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1004ABL01

Quant Time: Oct 05 04:39:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Sep 19 14:26:28 2022
Response via : Initial Calibration

