

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100422\
 Data File : VL039686.D
 Acq On : 4 Oct 2022 18:28
 Operator : SY/AP
 Sample : N4885-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1DL

Quant Time: Oct 05 04:41:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Sep 19 14:26:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.208	49	1032787	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.668	114	2759798	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.475	117	2506980	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.786	95	1793954	9.586	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.900%
Target Compounds						
					Qvalue	
3) Chlorodifluoromethane	2.703	51	61574	0.502	ppbv	96
15) Acetone	3.507	43	63488	0.552	ppbv #	87
20) Methylene Chloride	3.957	84	49525	0.932	ppbv	90
26) Hexane	5.230	57	57772	0.543	ppbv #	37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100422\
Data File : VL039686.D
Acq On : 4 Oct 2022 18:28
Operator : SY/AP
Sample : N4885-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1DL

Quant Time: Oct 05 04:41:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Sep 19 14:26:28 2022
Response via : Initial Calibration

