

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100422\
 Data File : VL039688.D
 Acq On : 4 Oct 2022 19:43
 Operator : SY/AP
 Sample : N4885-01DL2 80X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1DL2

Quant Time: Oct 05 04:42:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Sep 19 14:26:28 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/07/2022
 Supervised By :Mahesh Dadoda 10/07/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	1022319	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.668	114	2683450	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.474	117	2380180	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.783	95	1757053	9.889	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.900%
Target Compounds						
20) Methylene Chloride	3.954	84	14321m	0.272	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100422\
Data File : VL039688.D
Acq On : 4 Oct 2022 19:43
Operator : SY/AP
Sample : N4885-01DL2 80X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1DL2

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/07/2022
Supervised By :Mahesh Dadoda 10/07/2022

Quant Time: Oct 05 04:42:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Sep 19 14:26:28 2022
Response via : Initial Calibration

