

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100422\
 Data File : VL039692.D
 Acq On : 4 Oct 2022 22:17
 Operator : SY/AP
 Sample : N4955-13 LOD 0.40
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT4-2022

Quant Time: Oct 05 04:43:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Sep 19 14:26:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	964182	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.667	114	2421632	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.478	117	2151594	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.783	95	1566238	9.751	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.500%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.754	85	6149	0.055	ppbv	89
3) Chlorodifluoromethane	2.706	51	6739	0.059	ppbv #	85
15) Acetone	3.497	43	34530	0.322	ppbv	90
20) Methylene Chloride	3.960	84	108344	2.184	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100422\
Data File : VL039692.D
Acq On : 4 Oct 2022 22:17
Operator : SY/AP
Sample : N4955-13 LOD 0.40
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT4-2022

Quant Time: Oct 05 04:43:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Sep 19 14:26:28 2022
Response via : Initial Calibration

