

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100423\
 Data File : VL040749.D
 Acq On : 4 Oct 2023 20:53
 Operator : SY/MD
 Sample : 04699-13 LOD 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT4-2023

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/05/2023
 Supervised By :Mahesh Dadoda 10/05/2023

Quant Time: Oct 05 04:29:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091423AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2023
 QLast Update : Tue Sep 19 05:12:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	735863	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	1947839	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.449	117	1618946	10.000	ppbv	-0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.757	95	1272250	10.171	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.700%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.928	62	2616m	0.048	ppbv	
32) 1,1,1-Trichloroethane	5.996	97	5488m	0.042	ppbv	
35) Carbon Tetrachloride	6.475	117	5482m	0.041	ppbv	
38) Trichloroethene	7.288	130	2523m	0.046	ppbv	
52) Tetrachloroethene	10.603	164	1582m	0.034	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100423\
Data File : VL040749.D
Acq On : 4 Oct 2023 20:53
Operator : SY/MD
Sample : 04699-13 LOD 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT4-2023

Quant Time: Oct 05 04:29:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091423AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 19 05:12:43 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/05/2023
Supervised By :Mahesh Dadoda 10/05/2023

