

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100521\
Data File : VL037768.D
Acq On : 5 Oct 2021 10:50
Operator : SY/AP
Sample : VL1005ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1005ABL01

Quant Time: Oct 06 05:35:16 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Oct 04 17:43:54 2021
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1379046	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	3910229	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.06	117	3302855	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2294571	9.93	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.30%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100521\
Data File : VL037768.D
Acq On : 5 Oct 2021 10:50
Operator : SY/AP
Sample : VL1005ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1005ABL01

Quant Time: Oct 06 05:35:16 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 04 17:43:54 2021
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

