

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100521\
Data File : VL037768.D
Acq On : 5 Oct 2021 10:50
Operator : SY/AP
Sample : VL1005ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1005ABL01

Quant Time: Oct 06 05:35:16 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 04 17:43:54 2021
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1379046	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	3910229	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.06	117	3302855	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2294571	9.93	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.30%


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100521\
Data File : VL037768.D
Acq On : 5 Oct 2021 10:50
Operator : SY/AP
Sample : VL1005ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1005ABL01

Quant Time: Oct 06 05:35:16 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 04 17:43:54 2021
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.65 min Scan# 891

Delta R.T. -0.02 min

Lab File: VL037768.D

Acq: 5 Oct 2021 10:50

Instrument :

MSVOA_L

ClientSampleId :

VL1005ABL01


Tgt Ion: 49 Resp: 1379046

Ion Ratio Lower Upper

49 100

128 52.5 27.1 81.3




130 67.1 33.2 99.6



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): VL0

Ion 130.00 (129.70 to 130.70): VL0



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.16 min Scan# 1359

Delta R.T. -0.02 min

Lab File: VL037768.D

Acq: 5 Oct 2021 10:50


Tgt Ion: 114 Resp: 3910229

Ion Ratio Lower Upper

114 100

63 24.0 19.4 29.2



88 17.8 14.4 21.6




Abundance Ion 114.10 (113.80 to 114.80): VL0

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.06 min Scan# 2885
Delta R.T. -0.03 min
Lab File: VL037768.D
Acq: 5 Oct 2021 10:50

Instrument :
MSVOA_L
ClientSampleId :
VL1005ABL01

Tgt Ion: 117 Resp: 3302855
Ion Ratio Lower Upper
117 100
82 65.6 52.2 78.2
119 34.4 26.0 39.0



#68
1-Bromo-4-Fluorobenzene
Concen: 9.934 ppbv
RT: 14.39 min Scan# 3610
Delta R.T. -0.02 min
Lab File: VL037768.D
Acq: 5 Oct 2021 10:50

Tgt Ion: 95 Resp: 2294571
Ion Ratio Lower Upper
95 100
174 73.8 57.8 86.6
175 5.6 4.7 7.1

