

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100719\
 Data File : VL034234.D
 Acq On : 7 Oct 2019 22:56
 Operator : SY/AP
 Sample : K5225-04DL 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA2DL

Manual Integrations
 APPROVED

MMDadoda
 10/11/2019 5:05:29 PM

Quant Time: Oct 09 03:45:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
 QLast Update : Wed Sep 18 12:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	1111386	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.21	114	1826308	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.16	117	1856313	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.49	95	1664423	10.32	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chlorodifluoromethane	2.97	51	145935	1.120	ppbv	97
13) Ethanol	3.60	45	263031	16.895	ppbv	95
15) Acetone	3.85	43	235422	1.580	ppbv	99
19) Isopropyl Alcohol	3.97	45	62970m	0.687	ppbv	
34) 2-Butanone	5.27	43	33623m	0.180	ppbv	
76) 1,4-Dichlorobenzene	17.22	146	2380390	15.069	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100719\
Data File : VL034234.D
Acq On : 7 Oct 2019 22:56
Operator : SY/AP
Sample : K5225-04DL 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA2DL

Manual Integrations
APPROVED

MMDadoda
10/11/2019 5:05:29 PM

Quant Time: Oct 09 03:45:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep
QLast Update : Wed Sep 18 12:53:31 2019
Response via : Initial Calibration

