

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100719\
 Data File : VL034248.D
 Acq On : 8 Oct 2019 11:12
 Operator : SY/AP
 Sample : QC100719 CAN 10283
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC100719 CAN 10283

Quant Time: Oct 09 05:11:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
 QLast Update : Wed Sep 18 12:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	894074	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.21	114	1489181	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.15	117	1460347	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.49	95	1317411	10.38	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100719\
Data File : VL034248.D
Acq On : 8 Oct 2019 11:12
Operator : SY/AP
Sample : QC100719 CAN 10283
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC100719 CAN 10283

Quant Time: Oct 09 05:11:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
QLast Update : Wed Sep 18 12:53:31 2019
Response via : Initial Calibration

