

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100819\
Data File : VL034255.D
Acq On : 8 Oct 2019 19:13
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Oct 09 05:45:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
QLast Update : Wed Sep 18 12:53:31 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	1051232	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.21	114	1737438	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.15	117	1683010	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.49	95	1514667	10.36	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.60%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100819\
Data File : VL034255.D
Acq On : 8 Oct 2019 19:13
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Oct 09 05:45:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
QLast Update : Wed Sep 18 12:53:31 2019
Response via : Initial Calibration

